



SEC. X-X

NOTES:

1. DIMENSIONS GIVEN ARE FOR ENDEVCO 7264A-2000 AND 7264B-2000 ACCELEROMETERS ONLY.
2. ALL ACCELEROMETER MOUNTS ARE INTERCHANGEABLE.
3. POSITION ACCURACY OF BLOCKS IN HEAD  $\pm 0.005$ .
4. POSITION ACCURACY OF SEISMIC MASS WITHIN ACCELEROMETER  $\pm 0.01$  (PER ENDEVCO SPEC.).
5. MODEL 1716 LOAD CELL, ACCELEROMETERS, ACCELEROMETER ATTACHMENT SCREWS, HEAD SKIN, AND CAP SKIN ARE NOT INCLUDED.

NEXT ASSY	QTY				
PART USED ON					
		A	WAS C-4310	120CT99	
		SYM	REVISION	DATE	

ACCELEROMETER HEAD MOUNT FOR ENDEVCO 7264A & B FOR 5TH FEMALE			
SCALE:	FULL	DRAWN BY:	P. DEPINET
DATE:	04JAN99	MODEL:	4310
UNLESS OTHERWISE SPECIFIED			
TWO PLACE DECIMALS $\pm .010$		SURFACE FINISH 125 $\sqrt{R}$	
THREE PLACE DECIMALS $\pm .005$		REMOVE ALL BURRS & SHARP EDGES	
ANGLES $\pm 0' 30"$ DIMENSION LIMITS HELD AFTER PLATING			
<b>ROBERT A. DENTON, INC.</b> ROCHESTER HILLS, MICHIGAN			DRAWING NUMBER: <b>C-4310-D</b>